


Rotor, compatible, KaVo 9000/L. Myonic ceramic torlon bearings



Compatible rotor for KaVo Master Torque M9000L high speed handpiece, supplied with Myonic torlon ceramic bearings.

Available

SKU: HKC9000

Categories: 9000

Description

Compatible rotor for KaVo Master Torque M9000L, replacement part for dental high speed handpiece, supplied with Myonic torlon ceramic bearings.

Handpiece compatibility

- KaVo Master Torque M9000L

Referal KaVo rotor code: **2.000.2266**

Technical features

- Supplied with Myonic torlon ceramic bearings
- Dynamically balanced rotor
- Serial number identification
- Bur retention >40 Newton
- 12-month warranty
- Compliant with EN ISO 13485, EN ISO 10993-1, EN ISO 17665-1, EN ISO 5349 standards